

- [RSS Feeds](#)
- [Bookmarks](#)
- [Comments](#)

Researchers develop a halometer that tests alterations in night vision

Featured In: [International News](#)

By EurekAlert

Thursday, March 24, 2011

[Get daily Medical Design industry top stories and headlines - Sign up now!](#)



Email


Print

0
tweets

tweet

Share

0



This press release is available in [French](#) and [Spanish](#).

Researchers from the Department of Optics of the [University of Granada](#), belonging to the Laboratorio de Ciencias de la Visión y Aplicaciones, have developed a programme for testing alterations in night vision, and the tool required to implement it, which has been named "halometer".

This instrument consists on a software named Software Halo v1.0, and a computer where the mouse is used as a response button, and a chin cup with a forehead holder to fix the observer's position. Software Halo v1.0 was initially presented as freeware available on LabVisGra's website at <http://www.ugr.es/local/labvisgr> and at the University of Granada's institutional repository at <http://digibug.ugr.es/handle/10481/5478>. Nevertheless, it was recently released by the Free-Software Bureau of the [University of Granada](#) so, it has become a scientific freeware application.



This software was developed by a group coordinated by **Rosario González Anera**, receiving technical support of the Granadian company Seven Solutions and of an ophthalmological clinic set in Madrid and named Novovision. The research group was integrated by researchers **José Juan Castro**, **Carolina Ortiz** and **Aixa Alarcón**, and receives fundings from the regional government of Economy, Innovation and Science of the Junta de Andalucía.

Low-Light Conditions

Before the test, researchers fix the observer's position with a chin cup and a forehead holder at an adequate

distance from the monitor where the test will be displayed and considering the angular diameter that light stimuli must have. The test is performed under low-light conditions (in a dark room) and, before the test, patients are given some minutes to adapt to darkness.

This test involves showing a bright light stimulus to the subject against a dark background and, progressively showing different peripheral light stimuli with lower brightness around the central (Figure 2), at random, in different positions and at different distances. Subjects' task is to press the button of a mouse every time they detect a peripheral stimulus. Once the test ends, the software makes an estimation of a parameter named "Visual Disturbance Index" that, together with the results graphic, determines the type and intensity of visual disturbance.

This index scores from 0 to 1, so the higher is this index, the stronger is the influence of the halo or any other visual disturbance perceived by the observer. Therefore, the lower is the subject's ability to detect the peripheral stimuli displayed around the main stimulus that causes such disturbances.

One of the advantages of the halometer developed at the University of Granada is that it allows to configure such test parameters as:

- Special parameters: the radius of the central stimulus and the peripheral stimuli can be modified. Furthermore, the number of semi-major axes where the peripheral stimuli will be displayed can be fixed, as well as the maximum radius maximum distance between these stimuli-, and the number of stimuli per semi-major axis.
- Temporary Parameters: the software allows to set some exposure time for adaptation to darkness and to the brightness of the central stimulus. Furthermore, others parameters as the exposure time to the stimulus and the time elapsed between a stimulus and the next one can also be set.
- The software also allows to set the color and brightness of central and peripheral stimuli, and the weight and number of times each peripheral stimulus is shown.

This test has been tried with different observer groups consisting of people operated on cataract or submitted to refractory eye surgery, with normal observers and patients with different eye pathologies as AMD (Age-related Macular Degeneration) and unilateral keratitis. The results obtained showed that, in the case of patients with keratitis, the visual disturbance index for the healthy eye was found to be lower than that for the unhealthy eye, and even lower under binocular vision conditions. However, visual disturbance indexes proved to be higher in the case of patients operated on cataracts, showing a higher influence of halos and, therefore, lower detection capacity. Patients operated with LASIK refractory eye surgery also presented high vision disturbance indexes in comparison to normal observers.

The results obtained with this halometer and the new software have been presented in different national and international conferences as the IX Reunión Nacional de Óptica, Galice and the Wavefront Confress 2009. furthermore, a research article has been published in the international journal Applied Optics describing the results obtained with patients operated with refractory eye surgery. It has also been recently admitted for publication in the prestigious scientific journal *Journal of Biomedical Optics*, where the results obtained with patients with different eye pathologies are described.

[SOURCE](#)

Join the Discussion

Rate Article:  Average 0 out of 5
[register](#) or [log in](#) to comment on this article!

0 COMMENTS

ADD COMMENT